



Disposable Analog Forehead Temperature Trend Monitors

Instructions for Use

How to Read

1. Remove backing liner and apply monitor to center of forehead, just above the eyebrows. Skin surface should be clean and dry.
2. Read the monitor when yellow line stops moving (generally 15 seconds), and while it is still on the forehead.
3. Center of yellow line indicates the correct temperature

Intended Use

- Measure human forehead surface temperature and monitor temperature trends

Accuracy

- $\pm 1^{\circ}\text{F}$ (0.6°C)

Storage & Handling

- Store at 60 to 104°F in 15 to 95% RH
- Optimal storage at 60 to 80°F in 40 to 60% RH

Precautions

- Environmental temperatures might affect the performance of this product
- Product should be used indoors at normal room temperature
- Avoid exposure to UV light
- Discard after use

Liquid Crystal Forehead Temperature Trend Monitors are designed to be unbreakable noninvasive, quick reading devices that reliably indicate trends in core body temperature by measuring the surface temperature of the forehead. Liquid Crystal Forehead Temperature Trend Monitors indirectly measure core body temperature by measuring the forehead temperature, whereas clinical thermometers (oral/rectal) directly measure core body temperature. Unusual Liquid Crystal Forehead Temperature Trend Monitor readings should be verified by a clinical thermometer.

For external use only · Conforms to ASTM E-1061-01 · Not made with natural rubber latex

LCR Hallcrest, 1911 Pickwick Lane, Glenview IL 60026 USA

Hallcrest